



South African Technical Auditing Services

Certificate of Calibration

No 110/07305/21a

Galvaglow (Pty) Ltd
Attention : Mr P Chidemo
PO Box 1199
Krugersdorp
1740

Our Ref: 110/07305/21a
Enquiries K Ebersohn
Tel 083 632 3404
Date 23 January 2021

EQUIPMENT CALIBRATED

Description: Elcometer Serial No: LK 07305

METHOD OF CALIBRATION

The instrument was calibrated by means of certified calibrated shims - serial number B1-B3 traceable to national standards.

RESULTS

Nominal thickness of shims, μm	Measurement, μm
52,0 \pm 2,0	52
126,0 \pm 2,0	127
250,0 \pm 4,0	249

NOTE: The instrument was calibrated using a similar base material type and thickness to that used in practice.

Date of calibration: 21 January 2021
Date of expiry: 20 January 2022

Kevin Ebersohn
SATAS

SOUTH AFRICAN TECHNICAL AUDITING SERVICES (Co No 2002/015355/07)

Directors JA Stears Cell 083 632 3416 Fax 012 345 6646 E mail stearsa@satas.co.za
DV Goodall Cell 083 632 3260 Fax 031 708 3635 Email goodalld@satas.co.za
PL Posthumus Cell 083 632 4884 Fax 013751 2159 Email posthumusp@satas.co.za